



Frequency Evaluation of UTC(NIST) by NIST-Yb1 for the period MJD 59974 to 59999

I. Results

Period	y(UTC(NIST) - Yb1)	u _A	u _B	u _{A/Lab}	u _{B/Lab}
(MJD)	(10 ⁻¹⁶)				
59974- 59999	9.7	<0.1	0.062	4.0	0.3

u _A , Type A uncertainty (10 ⁻¹⁶)			
Yb stability	<0.1		
Total	<0.1		

u _B , Type B Uncertainty (10 ⁻¹⁶)		
Yb total systematic	0.014	
Gravitational redshift	0.06	
Total	0.062	

u _{A/Lab} , local link Type A uncertainty (10 ⁻¹⁶)				
Dead time	3.7			
Yb-Maser comparison	1.4			
Time scale measurement	<0.1			
Total	4.0			

u _{B/Lab} , local link Type B uncertainty (10 ⁻¹⁶)			
Frequency comb + counting	0.3		
Microwave transmission	<0.1		
Total	0.3		

II. NIST-Yb1 operation

During the indicated period, NIST-Yb1 and an optical frequency comb were operated intermittently with a combined uptime of 5.0%. The measured frequency difference assumes the Yb absolute frequency equal to the most recently published CCTF recommendation: 518,295,836,590,863.63 Hz [1]. More details on NIST-Yb1 clock operation and its typical uncertainty budget can be found in Ref. 2 and 3.

Effect	Shift (10 ⁻¹⁸)	Uncertainty(10 ⁻¹⁸)	
Background gas collisions	-5.5	0.5	
Spin polarization	0	<0.3	
Cold collisions	-0.21	0.07	
Doppler	0	<0.02	
Blackbody radiation	-2,361.2	0.9	
Lattice light (model)	0	0.3	
Travelling wave contamination	0	<0.1	
Lattice light (experimental)	-1.5	0.8	
Second-order Zeeman	-118.1	0.2	
DC Stark	0	< 0.07	
Probe Stark	0.02	0.01	
Line pulling	0	<0.1	
Tunneling	0	< 0.001	
Servo error	0.03	0.05	
Optical frequency synthesis	0	<0.1	
Yb1 Total	-2,486.5	1.4	
Grav. redshift from geoid [4]	180,819	6	
Yb + gravitational redshift	178,333	6.2	

Typical NIST-Yb1 systematic biases and uncertainties

III. Frequency measurement

The frequency measurement was carried out with an optical frequency comb that was phase-locked to NIST-Yb1, and the resulting comb frequencies were subsequently counted relative to a hydrogen maser, 412014. For this analysis, one-second gated counting data (measured with a software-defined-radio-based frequency counter) were binned into twelve minute intervals, and related to internal NIST timescales. A final average value was calculated over the indicated period. A breakdown of the Type A and Type B uncertainties for this measurement are listed in the results section. Dead time uncertainty associated with the less-than-unit uptime of the NIST-Yb1 measurement during the indicated period is calculated following the method of [5] and as outlined in [3]. The reported frequency offset, y(UTC(NIST)-Yb1), is computed with NIST-Yb1 frequency corrections from the geoid [4,2].

[1] "Recommended values of standard frequencies for applications including the practical realization of the metre and secondary representations of the definition of the second," BIPM publication, approved by CCTF March 2021, <u>https://www.bipm.org/documents/20126/69375133/171Yb_518THz_2021.pdf/283dca33-4dac-f309-671e-577af2a62fc1</u>.

[2] W. McGrew, et al., "Atomic clock performance enabling geodesy below the centimetre level," Nature **564**, 87–90 (2018).

[3] W. McGrew, et al., "Towards Adoption of an Optical Second: Verifying Optical Clocks at the SI Limit," Optica **6**, 448-454 (2019).

[4] N. K. Pavlis and M. A. Weiss, "A re-evaluation of the relativistic redshift on frequency standards at NIST, Boulder, Colorado, USA," Metrologia **54**, 535-548 (2017).

[5] D.-H. Yu, M. Weiss, and T. E. Parker, "Uncertainty of a frequency comparison with distributed dead time and measurement interval offset," Metrologia 44, 91–96 (2007).