# Evaluation of PTB primary caesium fountain frequency standard CSF1 between MJD 58904 - MJD 58939

PTB's primary caesium fountain frequency standard CSF1 was operated between MJD 58904, 0:00 UTC and MJD 58939, 0:00 UTC. Frequency comparisons were made with respect to PTB hydrogen maser H9, BIPM code 1400509.

The relative frequency instability of the relative frequency differences y(CSF1-H9) was  $12.1\times10^{-14}\cdot(\tau/s)^{-1/2}$  during the 35 days. The actual measurement time amounts to 89.9% of the 35 × 24 hours. This results in a statistical uncertainty  $u_A = 0.07\times10^{-15}$ , assuming that white frequency noise is the dominant noise source.

For the uncertainty due to the clock link  $u_{Lab} = 0.06 \times 10^{-15}$  is obtained by taking into account the actual measurement time. Finally, the estimated uncertainty for the link to TAI for 35 days is  $u_{TAI} = 0.06 \times 10^{-15}$ .

Frequency corrections for the following effects were applied to the raw data:

- Zeeman effect (magnetic field along the atoms' trajectory)
- black body effect (thermal radiation along the atoms' trajectory)
- relativistic redshift and relativistic Doppler effect
- cold collisions effect
- distributed cavity phase effect
- microwave lensing effect

The CSF1 standard uncertainty  $u_{\rm B}$  is estimated as 3.1 ×10<sup>-16</sup> (1  $_{\rm O}$ ) for the relevant period [1].

### Table of results of CSF1 compared to hydrogen maser H9 (1400509)

Interval of evaluation MJD 58904, 0:00 UTC – MJD 58939, 0:00 UTC

Fractional dead time 10.1%

Resulting frequency difference  $y(CSF1 - H9) = 8.14 \times 10^{-15}$ 

Type A uncertainty  $u_A$  (1  $\sigma$ ) 0.07  $\times$  10<sup>-15</sup>

Type B uncertainty  $u_{\rm B}$  (1  $\sigma$ ) 0.31  $\times$  10<sup>-15</sup>

Link to clock  $u_{Lab}$  (1  $\sigma$ )  $0.06 \times 10^{-15}$ 

Link to TAI  $u_{TAI}$  (1  $\sigma$ ) 0.06 × 10<sup>-15</sup> (35 days)

Combined uncertainty (1  $\sigma$ ) 0.33  $\times$  10<sup>-15</sup>

#### Type A (statistical) uncertainty of CSF1

For the microwave synthesis an optically stabilized microwave oscillator is utilized, which is locked to a hydrogen maser in the long-term [2]. The frequency instability  $12.1\times10^{-14}~(\tau/s)^{-1/2}$  of the measured relative frequency differences y(CSF1 – Hmaser) is obtained for the combination of low and high density operation and gives the statistical measurement uncertainty  $u_A$  [1].

The optically stabilized microwave system was available during about 89% of the TAI measurement interval. Alternatively a quartz-based frequency synthesis system was employed.

#### Type B (systematic) uncertainty of CSF1

In the table below we report the type B uncertainty evaluation results valid for the evaluation at hand. Detailed descriptions of the systematic uncertainty contributions of CSF1 have been published elsewhere [1].

At the 26<sup>th</sup> CGPM in November 2018, TAI has been newly defined (Resolution 2). As a result the relativistic redshift of a clock contributing to TAI is to be computed with respect to the conventionally adopted equipotential  $W_0 = 62\,636\,856.0\,\text{m}^2\text{s}^{-2}$  of the Earth's gravity potential. The differentiation relating to the uncertainty of the relativistic redshift for the case of TAI contributions of the PTB fountain clocks in [1] is therefore no longer needed, so that a relativistic redshift uncertainty of  $0.02 \times 10^{-16}$  [1] is attributed now and in the future.

# Frequency shifts, corrections and type B uncertainties of CSF1 (parts in 10<sup>16</sup>):

Frequency shift	Correction	Uncertainty
Quadratic Zeeman shift	- 1079.09	0.10
Blackbody radiation shift	166.66	0.81
Relativistic redshift and Doppler effect	- 85.56	0.02
Collisional shift	-17.2	2.8
Distributed cavity phase shift	- 0.04	0.93
Microwave lensing	-0.4	0.2
AC Stark shift (light shift)		0.01
Rabi and Ramsey pulling		0.013
Microwave leakage		0.01
Electronics		0.1
Background gas collisions		0.4
Total type B uncertainty		3.1

## References

[1] S. Weyers, V. Gerginov, M. Kazda, J. Rahm, B. Lipphardt, G. Dobrev and K. Gibble, Metrologia **55**, pp. 789–805 (2018), <a href="https://doi.org/10.1088/1681-7575/aae008">https://doi.org/10.1088/1681-7575/aae008</a>

[2] B. Lipphardt, V. Gerginov, S, Weyers, IEEE Transactions on Ultrasonics, Ferroelectrics, and Frequency Control **64**, pp. 761–766 (2017), <a href="https://ieeexplore.ieee.org/document/7807353">https://ieeexplore.ieee.org/document/7807353</a>