

Evaluation of the frequency of H-maser 1401708 by primary frequency standard NPL-CsF2

National Physical Laboratory

1st December 2020

The primary frequency standard NPL-CsF2 was used to measure the frequency of hydrogen maser HM2, identified by clock code 1401708, over an evaluation period in October and November 2020. Maser 1401708 is a physical realisation of UTC(NPL). The evaluation was performed by measuring the mean frequency difference over the reporting period.

No changes to NPL-CsF2 or its associated operating protocols have been introduced since the previously reported evaluation, and a breakdown of the systematic uncertainties is reproduced in table Table 1. Note that the uncertainty contribution from cold collisions varies and the value listed here is purely indicative. A specific value for the given measurement period is presented together with the measurement results.

	uncertainty / 10^{-16}
Second order Zeeman	0.8
Blackbody radiation	1.0
AC Stark (lasers)	0.1
Microwave spectrum	0.1
Gravity	0.5
Cold collisions	0.4 [†]
Background gas collisions	0.3
Rabi, Ramsey pulling	0.1
Cavity phase (distributed)	1.0
Cavity phase (dynamic)	0.1
Cavity pulling	0.6
Microwave leakage	0.6
Microwave lensing	0.3
2 nd -order Doppler	0.1
Total u_B (1σ)	2.0

Table 1: Results of the most recent accuracy evaluation of NPL-CsF2.

[†] The exemplary value here corresponds to the type B uncertainty contribution for a ratio of high to low atom density of 8 and a measured frequency difference between the two of below 2.5×10^{-15} .

Measurement results

Results of the frequency measurement are listed in table Table 2 below. Frequency biases are given for information only and represent the mean values of the biases over the measurement interval. The listed fractional frequency difference $y(\text{CsF2-HM2})$ is a value corrected for these biases. The total uncertainty u_{total} is defined as:

$$(u_{\text{total}})^2 = (u_A)^2 + (u_B)^2 + (u_{A/\text{lab}})^2 + (u_{B/\text{lab}})^2$$

Maser 1401708 is steered on a regular basis and the measurements listed are a weighted average of the measurements throughout each period.

		27 Oct 2020 – 26 Nov 2020
Period start	MJD	59149
Period end	MJD	59179
Duration	days	30
Measurement uptime	%	96.8
Biases:	$\times 10^{-15}$	
cold collisions		0.19
2 nd order Zeeman		247.57
BBR shift		-16.34
gravity		1.30
microwave lensing		0.06
DCP		0.02
$y(\text{CsF2-HM2})$	$\times 10^{-15}$	-0.07
u_A	$\times 10^{-15}$	0.14
u_B	$\times 10^{-15}$	0.20
$u_{A/\text{lab}}$	$\times 10^{-15}$	0.04
$u_{B/\text{lab}}$	$\times 10^{-15}$	0.00
u_{total}	$\times 10^{-15}$	0.25

Table 2: Results of the evaluation of the frequency of H-maser 1401708 by primary frequency standard NPL-CsF2.