

**FREQUENCY COMPARISON (H\_MASER 140 0889) - (LNE-SYRTE-FO2)**  
**For the period MJD 57934 to MJD 57964**

The primary frequency standard LNE-SYRTE-FO2 has been compared to the hydrogen Maser 140 0889 of the laboratory, during two measurement campaigns between MJD 57934 and 57964 (30<sup>th</sup> June 2017 – 30<sup>th</sup> July 2017). The behaviour of the maser was perturbed due to problems with the air conditioning system of the clock room. Therefore, the collected data have been separated into two intervals of 10 and 15 days respectively. The fountain operation covers ~ 89.4% and ~ 95.5% of the two measurement durations.

The mean frequency difference at the middle date of each period is given in the following table:

Period (MJD)	Date of the estimation	$y(\text{HMaser}_{140\ 0889} - \text{FO2})$	$u_B$	$u_A$	$u_{\text{link} / \text{maser}}$
57934 – 57944	57939	-3802.6	2.4	6.0	1.1
57949 – 57964	57956.4	-3835.6	2.4	2.4	1.0

*Table 1: Results of the comparison in  $1 \times 10^{-16}$ .*

The FO2 fountain was operated in the same mode during all the period: the interrogating signal is based on the down conversion to 9.192 GHz of a 11.98 GHz signal provided by a cryogenic oscillator phase locked to the maser 140 0889. A synthesizer is used to lock the microwave signal on the atomic resonance. The frequency difference between this maser and the fountain is deduced from the average correction applied to the synthesizer.

### Average value and statistical uncertainty

The frequency data are averaged over 0.2 day intervals. We then calculate with the average points, a linear unweighted fit to determine the average frequencies given in Table 1. The statistical uncertainties are estimated with the Allan variance of the frequency residuals, after removing the drift. We estimate conservative statistical uncertainties  $u_A = 6.0 \times 10^{-16}$  and  $u_A = 2.4 \times 10^{-16}$  for the two segments respectively.

We verified the results by applying a second method. We calculated the accumulated phase by integrating the data points, assuming a constant frequency during each segment, and during the dead times of the fountain operation. The average frequency is then obtained by dividing the total accumulated phase by the calibration period duration. The processing has been performed with segments of 0.01, 0.1 and 1 day durations. The results are in agreement with the values given in Table 1 within  $0.7 \times 10^{-16}$  and  $0.1 \times 10^{-16}$  for the two segments respectively, which is consistent with the estimations of the statistical uncertainties  $u_A$  and with the uncertainties due to the link.

## Accuracy

The frequency is corrected from the quadratic Zeeman, the Black Body radiation, the cold collisions and cavity pulling, the first order Doppler, the microwave lensing, and the red shift effects. Here the uncertainty in the cold collisions correction accounts for both a statistical uncertainty and a systematic uncertainty taken as 0.5% of the correction for high density measurements. The following table summarizes the budget of systematic effects and their associated uncertainties. The accuracy is the quadratic sum of all the systematic uncertainties.

	Correction (10 <sup>-16</sup> )	Uncertainty (10 <sup>-16</sup> )
Quadratic Zeeman effect	-1919.2	0.3
Black body radiation	168.4	0.6
Cold collisions and cavity pulling	96.4	1.2
Distributed cavity phase shift	-0.9	1.0
Microwave spectral purity&leakage	0	< 0.5
Ramsey & Rabi pulling	0	< 0.1
Microwave lensing	-0.7	0.7
Second order Doppler effect	0	< 0.1
Background gas collisions	0	<1.0
<b>Total</b>	<b>-1656.0</b>	<b>2.2</b>
Red shift	- 65.5	1.0
<b>Total with red shift</b>	<b>-1721.5</b>	<b>2.4</b>

Table 2: Budget of systematic effects and uncertainties for SYRTE-FO2 fountain for the MJD 57934 – 57964 period

$$u_B = 2.4 \times 10^{-16}$$

## Uncertainty of the link

The uncertainty of the link is the quadratic sum of 2 terms:

-A possible effect of phase fluctuations introduced by the cables that connect the primary standard to the Maser. It is estimated to be 10<sup>-16</sup>.

-The uncertainty due to the dead times of the frequency comparison.

To estimate this contribution, we use the comparison between the reference Maser and Maser 140 0809.

We calculate the time deviation of the normalized phase differences with the linear frequency drift removed. The uncertainty is given by:

$$\sigma_{y_{Dead\ Time}} = \frac{\sqrt{\sum_i \sigma_{x_i}^2}}{T}$$

where  $\sigma_{x_i}$  are the extrapolated TVar for each dead times. We applied the method to the dead times longer than 600 s and obtained stability degradations of  $0.4 \times 10^{-16}$  and  $0.2 \times 10^{-16}$  for the two segments, respectively.