

Table 6. Measurements of the duration of the TAI scale interval

TAI is a realization of coordinate time TT. The following tables give the fractional deviation d of the scale interval of TAI from that of TT (in practice the SI second on the geoid), i.e. the fractional frequency deviation of TAI with the opposite sign: $d = -y_{TAI}$.

In this table, d is obtained on the given periods of estimation by comparison of the TAI frequency with that of the individual primary frequency standards (PFS) IT-CSF1, NICT-CSF1, NIST-F1, NMIJ-F1, PTB-CS1, PTB-CS2, PTB-CSF1, SYRTE-FO1, SYRTE-FO2, SYRTE-FOM and SYRTE-JPO for the year 2008.

Previous calibrations are available in the successive annual reports of the BIPM Time Section volumes 1 to 18 and in the BIPM annual report on time activities volume 1 and 2.

Each comparison is provided with the following information:

u_A is the uncertainty originating in the instability of the PFS,

u_B is the combined uncertainty from systematic effects,

Ref(u_B) is a reference giving information on the stated value of u_B ,

$u_{link/lab}$ is the uncertainty in the link between the PFS and the clock participating to TAI, including the uncertainty due to dead-time,

$u_{link/TAI}$ is the uncertainty in the link to TAI, computed using the standard uncertainty of [UTC-UTC(k)],

u is the quadratic sum of all four uncertainty values.

In this table, a frequency over a time interval is defined as the ratio of the end-point phase difference to the duration of the interval.

The typical characteristics of the calibrations of the TAI frequency provided by the different primary standards over 2008 are indicated below. Reports of individual PFS evaluations may be found at <http://www.bipm.org> in the subdirectory named 'data'.

Primary Standard	Type /selection	Type B std. Uncertainty	Operation	Comparison with	Number/typical duration of comp.
IT-CSF1	Fountain	(0.5 to 0.7) $\times 10^{-15}$	Discontinuous	H maser	6 / 10 to 20 d
NICT-CSF1	Fountain	(0.8 to 1.5) $\times 10^{-15}$	Discontinuous	UTC(NICT)	2 / 10-15 d
NIST-F1	Fountain	0.3 $\times 10^{-15}$	Discontinuous	H maser	5 / 15 to 25 d
NMIJ-F1	Fountain	3.9 $\times 10^{-15}$	Discontinuous	H maser	7 / 15 to 25 d
PTB-CS1	Beam /Mag.	8 $\times 10^{-15}$	Continuous	TAI	10 / 30 d
PTB-CS2	Beam /Mag.	12 $\times 10^{-15}$	Continuous	TAI	12 / 30 d
PTB-CSF1	Fountain	0.9 $\times 10^{-15}$	Discontinuous	H maser	2 / 25 d
SYRTE-F01	Fountain	(0.4 to 0.6) $\times 10^{-15}$	Discontinuous	H maser	8 / 10 to 30 d
SYRTE-F02	Fountain	(0.4 to 0.6) $\times 10^{-15}$	Discontinuous	H maser	9 / 10 to 30 d
SYRTE-FOM	Fountain	(0.7 to 0.9) $\times 10^{-15}$	Discontinuous	H maser	6 / 10 to 30 d
SYRTE-JPO	Beam /Opt.	6.3 $\times 10^{-15}$	Discontinuous	H maser	12 / 10 to 30 d

More detailed information on the characteristics and operation of individual PFS may be found in the annexes supplied by the individual laboratories.

Table 6. (Cont.)

Standard	Period of estimation	d (10^{-15})	u_A (10^{-15})	u_B (10^{-15})	Ref(u_B)	$u_{\text{link/lab}}$ (10^{-15})	$u_{\text{link/TAI}}$ (10^{-15})	u (10^{-15})	Notes
IT-CSF1	54604 54614	1.7	1.1	0.5	[1]	0.4	0.9	1.5	
IT-CSF1	54614 54634	2.5	1.0	0.5		0.2	0.5	1.2	
IT-CSF1	54649 54669	4.5	0.9	0.5		0.2	0.5	1.1	
IT-CSF1	54709 54729	2.7	0.7	0.7		0.2	0.6	1.2	
IT-CSF1	54754 54769	1.6	0.7	0.4		0.2	0.6	1.0	
IT-CSF1	54774 54789	4.3	0.8	0.7		0.2	0.6	1.2	
NICT-CsF1	54534 54544	2.9	1.0	1.5	[2]	0.3	0.9	2.0	
NICT-CsF1	54709 54724	-0.1	1.0	0.8		0.3	0.9	1.6	
NIST-F1	54469 54494	3.0	0.3	0.3	[3]	0.3	0.4	0.6	
NIST-F1	54554 54569	4.9	0.3	0.3		0.5	0.6	0.9	
NIST-F1	54654 54674	4.1	0.3	0.3		0.3	0.5	0.7	
NIST-F1	54774 54799	6.1	0.1	0.3		0.1	0.4	0.5	
NIST-F1	54814 54829	4.1	0.4	0.3		0.5	0.6	0.9	
NMIJ-F1	54504 54519	3.2	0.9	3.9	[4]	0.4	0.9	4.1	
NMIJ-F1	54529 54549	1.3	0.8	3.9		0.3	0.7	4.0	
NMIJ-F1	54559 54584	-0.1	0.7	3.9		0.3	0.5	4.0	
NMIJ-F1	54594 54614	1.2	0.8	3.9		0.3	0.7	4.0	
NMIJ-F1	54679 54704	2.9	0.7	3.9		0.3	0.5	4.0	
NMIJ-F1	54714 54739	1.6	0.7	3.9		0.3	0.5	4.0	
NMIJ-F1	54774 54789	-0.1	0.9	3.9		0.3	0.9	4.1	
PTB-CS1	54464 54494	-6.7	5.0	8.0	[5]	0.0	0.1	9.4	(1)
PTB-CS1	54494 54524	-9.7	5.0	8.0		0.0	0.1	9.4	
PTB-CS1	54524 54554	3.7	5.0	8.0		0.0	0.1	9.4	
PTB-CS1	54554 54584	2.6	5.0	8.0		0.0	0.1	9.4	
PTB-CS1	54584 54614	5.0	5.0	8.0		0.0	0.1	9.4	
PTB-CS1	54614 54644	0.4	5.0	8.0		0.0	0.1	9.4	
PTB-CS1	54644 54674	-5.7	5.0	8.0		0.0	0.1	9.4	
PTB-CS1	54674 54709	-2.1	5.0	8.0		0.0	0.1	9.4	
PTB-CS1	54709 54739	-13.1	5.0	8.0		0.0	0.2	9.4	
PTB-CS1	54739 54769	-12.5	5.0	8.0		0.0	0.2	9.4	
PTB-CS2	54464 54494	1.7	3.0	12.0	[6]	0.0	0.1	12.4	(1)
PTB-CS2	54494 54524	7.5	3.0	12.0		0.0	0.1	12.4	
PTB-CS2	54524 54554	1.2	3.0	12.0		0.0	0.1	12.4	
PTB-CS2	54554 54584	7.2	3.0	12.0		0.0	0.1	12.4	
PTB-CS2	54584 54614	0.8	3.0	12.0		0.0	0.1	12.4	
PTB-CS2	54614 54644	-1.2	3.0	12.0		0.0	0.1	12.4	
PTB-CS2	54644 54674	-1.1	3.0	12.0		0.0	0.1	12.4	
PTB-CS2	54674 54709	-3.1	3.0	12.0		0.0	0.1	12.4	
PTB-CS2	54709 54739	0.9	3.0	12.0		0.0	0.2	12.4	
PTB-CS2	54739 54769	4.7	3.0	12.0		0.0	0.2	12.4	
PTB-CS2	54769 54799	4.1	3.0	12.0		0.0	0.1	12.4	
PTB-CS2	54799 54829	4.3	3.0	12.0		0.0	0.2	12.4	
PTB-CSF1	54644 54669	6.7	0.1	0.9	[7]	0.0	0.2	0.9	
PTB-CSF1	54679 54704	6.1	0.1	0.9		0.0	0.2	0.9	

Table 6. (Cont.)

Standard	Period of estimation	d (10^{-15})	u_A (10^{-15})	u_B (10^{-15})	Ref(u_B)	$u_{\text{link/lab}}$ (10^{-15})	$u_{\text{link/TAI}}$ (10^{-15})	u (10^{-15})	Notes
SYRTE-F01	54484 54494	2.4	0.4	0.4	[8]	0.1	0.9	1.1	
SYRTE-F01	54554 54584	3.8	0.3	0.6		0.1	0.3	0.8	
SYRTE-F01	54589 54599	3.9	0.7	0.6		0.1	0.9	1.3	
SYRTE-F01	54614 54644	3.0	0.4	0.4		0.3	0.3	0.7	
SYRTE-F01	54644 54669	2.5	0.3	0.4		0.1	0.4	0.7	
SYRTE-F01	54734 54754	4.8	0.2	0.5		0.2	0.5	0.7	
SYRTE-F01	54779 54794	3.8	0.3	0.4		0.1	0.6	0.8	
SYRTE-F01	54799 54829	4.6	0.6	0.5		0.1	0.3	0.8	
SYRTE-F02	54554 54584	4.4	0.9	0.4	[8]	0.1	0.3	1.1	
SYRTE-F02	54589 54614	5.2	0.2	0.4		0.3	0.4	0.6	
SYRTE-F02	54614 54644	3.9	0.3	0.4		0.2	0.3	0.6	
SYRTE-F02	54644 54669	2.7	0.3	0.4		0.1	0.4	0.7	
SYRTE-F02	54674 54694	3.9	0.3	0.4		0.2	0.5	0.7	
SYRTE-F02	54714 54729	3.1	0.6	0.4		0.2	0.6	1.0	
SYRTE-F02	54759 54769	3.6	0.3	0.5		0.2	0.9	1.1	
SYRTE-F02	54779 54794	4.5	0.2	0.6		0.1	0.6	0.9	
SYRTE-F02	54799 54829	4.9	0.6	0.5		0.1	0.3	0.8	
SYRTE-F0M	54464 54494	3.0	0.2	0.9	[9]	0.1	0.4	1.0	
SYRTE-F0M	54499 54524	3.4	0.2	0.7		0.1	0.4	0.8	
SYRTE-F0M	54544 54554	4.4	0.5	0.7		0.1	0.9	1.2	
SYRTE-F0M	54554 54584	3.5	0.3	0.7		0.1	0.3	0.8	
SYRTE-F0M	54589 54614	4.6	0.2	0.7		0.2	0.4	0.9	
SYRTE-F0M	54614 54639	2.8	0.2	0.7		0.1	0.4	0.8	
SYRTE-JP0	54484 54494	2.2	1.2	6.3	[10]	0.3	0.9	6.5	
SYRTE-JP0	54494 54524	2.9	1.0	6.3		0.3	0.3	6.4	
SYRTE-JP0	54544 54554	0.7	1.0	6.3		0.3	0.9	6.4	
SYRTE-JP0	54554 54584	-1.3	0.6	6.3		0.3	0.3	6.3	
SYRTE-JP0	54589 54614	1.3	0.7	6.3		0.3	0.4	6.4	
SYRTE-JP0	54614 54644	2.7	0.6	6.3		0.3	0.3	6.3	
SYRTE-JP0	54644 54674	2.5	0.5	6.3		0.3	0.3	6.3	
SYRTE-JP0	54674 54694	1.5	0.7	6.3		0.5	0.5	6.4	
SYRTE-JP0	54714 54734	-0.1	0.6	6.3		0.5	0.5	6.4	
SYRTE-JP0	54744 54769	4.0	0.9	6.3		0.3	0.4	6.4	
SYRTE-JP0	54769 54794	2.0	1.1	6.3		0.3	0.4	6.4	
SYRTE-JP0	54799 54814	6.7	1.3	6.3		0.3	0.6	6.5	

Notes:

(1) Continuously operating as a clock participating to TAI.

References:

- [1] Levi F. et al., *Metrologia* 43-6, 545, 2006.
[2] Kumagai M., et al., *Metrologia* 45-2, 139-148, 2008.
[3] Heavner T.P. et al., *Metrologia* 42, 411, 2005. Parker T.E. et al., *Metrologia* 42, 423, 2005.
[4] Kurosu T. et al., *IEEE Trans. IM-53*, 466, 2004.
[5] Bauch A. et al., *Metrologia* 35, 829, 1998; Bauch A., *Metrologia* 42, S43, 2005.
[6] Bauch A. et al., *IEEE Trans. IM-36*, 613, 1987; Bauch A., *Metrologia* 42, S43, 2005.
[7] Weyers S. et al., *Metrologia* 38-4, 343, 2001; Weyers S. et al., Proceedings of the 6th Symposium on Frequency Standards and Metrology, University of St Andrews, World Scientific pub., 64-71, 2001.
[8] Vian C. et al., *IEEE Trans. IM-54*, 833, 2005 ; Vian C. et al., *Proc 19th EFTF*, 53, 2005.
[9] Marion H. et al. *Phys. Rev. Lett.*, 90, 150801, 2003.
[10] Makdissi A. and de Clercq E., *Metrologia* 38-5, 409, 2001.