

**Table 6. Measurements of the duration of the TAI scale interval**(File available at <ftp://62.161.69.5/pub/tai/scale/UTAI/utai09.ar>)

TAI is a realization of coordinate time TT. The following tables give the fractional deviation  $d$  of the scale interval of TAI from that of TT (in practice the SI second on the geoid), i.e. the fractional frequency deviation of TAI with the opposite sign:  $d = -y_{\text{TAI}}$ .

In this table,  $d$  is obtained on the given periods of estimation by comparison of the TAI frequency with that of the individual primary frequency standards (PFS) IT-CSF1, KRISS-1 NICT-CSF1, NIST-F1, NMIJ-F1, PTB-CS1, PTB-CS2, PTB-CSF1, PTB-CSF2, SYRTE-FO1, SYRTE-FO2, SYRTE-FOM and SYRTE-JPO for the year 2009. Previous calibrations are available in the successive annual reports of the BIPM Time Section volumes 1 to 18 and in the BIPM annual report on time activities volume 1 to 3.

Each comparison is provided with the following information:

$u_A$  is the uncertainty originating in the instability of the PFS,

$u_B$  is the combined uncertainty from systematic effects,

$u_{\text{link/lab}}$  is the uncertainty in the link between the PFS and the clock participating to TAI, including the uncertainty due to dead-time,

$u_{\text{link/TAI}}$  is the uncertainty in the link to TAI, computed using the standard uncertainty of  $[UTC - UTC(k)]$ ,

$u$  is the quadratic sum of all four uncertainty values.

In this table, a frequency over a time interval is defined as the ratio of the end-point phase difference to the duration of the interval.

The typical characteristics of the calibrations of the TAI frequency provided by the different primary standards over 2009 are indicated below. Reports of individual PFS evaluations may be found at [ftp://62.161.69.5/pub/tai/data/PFS\\_reports](ftp://62.161.69.5/pub/tai/data/PFS_reports).  $u_B(\text{Ref})$  is a reference giving information on the stated value of  $u_B$ ,  $u_B(\text{Ref})$  is the  $u_B$  value stated in this reference. Note that the current  $u_B$  values are generally not the same as the peer reviewed values given in  $\text{Ref}(u_B)$ .

Primary Standard	Type /selection	Type B std. uncertainty	$u_B(\text{Ref})/10^{-15}$	$\text{Ref}(u_B)$	Comparison with	Number/typical duration of comp.
IT-CSF1	Fountain	$0.7 \times 10^{-15}$	0.5	[ 1]	H maser	2 / 20 d to 25 d
KRISS-1	Beam /Opt.	$9.5 \times 10^{-15}$	9.5	[ 2]	H maser	3 / 10 d to 20 d
NICT-CSF1	Fountain	$(0.9 \text{ to } 1.3) \times 10^{-15}$	1.9	[ 3]	UTC(NICT)	2 / 10 d to 25 d
NIST-F1	Fountain	$(0.3 \text{ to } 0.6) \times 10^{-15}$	0.35	[ 4]	H maser	9 / 10 d to 20 d
NMIJ-F1	Fountain	$3.9 \times 10^{-15}$	3.9	[ 5]	H maser	7 / 20 d to 30 d
PTB-CS1	Beam /Mag.	$8 \times 10^{-15}$	8.	[ 6]	TAI	12 / 30 d
PTB-CS2	Beam /Mag.	$12 \times 10^{-15}$	12.	[ 7]	TAI	12 / 30 d
PTB-CSF1	Fountain	$(0.8 \text{ to } 0.9) \times 10^{-15}$	1.4	[ 8]	H maser	2 / 20 d to 30 d
PTB-CSF2	Fountain	$(0.7 \text{ to } 1.4) \times 10^{-15}$	0.8	[ 9]	H maser	6 / 10 d to 20 d
SYRTE-F01	Fountain	$(0.4 \text{ to } 0.5) \times 10^{-15}$	0.72	[10]	H maser	10 / 10 d to 30 d
SYRTE-F02	Fountain	$(0.4 \text{ to } 0.5) \times 10^{-15}$	0.65	[10]	H maser	8 / 10 d to 30 d
SYRTE-FOM	Fountain	$0.7 \times 10^{-15}$	0.80	[11]	H maser	4 / 15 d to 30 d
SYRTE-JPO	Beam /Opt.	$6.3 \times 10^{-15}$	6.3	[12]	H maser	13 / 5 d to 30 d

More detailed information on the characteristics and operation of individual PFS may be found in the annexes supplied by the individual laboratories.

Table 6. (Cont.)

Standard	Period of estimation		$d/10^{-15}$	$u_A/10^{-15}$	$u_B/10^{-15}$	$u_{\text{link/lab}}/10^{-15}$	$u_{\text{link/TAI}}/10^{-15}$	$u/10^{-15}$	Notes
IT-CsF1	55139	55159	3.99	0.40	0.70	0.22	0.47	0.96	
IT-CsF1	55164	55189	3.57	0.70	0.70	0.40	0.38	1.13	
KRISS-1	54654	54669	-4.8	3.0	9.5	1.0	0.9	10.0	
KRISS-1	54699	54709	-6.9	4.0	9.5	1.0	2.6	10.7	
KRISS-1	54719	54739	2.3	2.0	9.5	1.0	1.4	9.9	
NICT-CsF1	55134	55144	1.73	1.00	1.30	0.30	0.53	1.75	
NICT-CsF1	55159	55184	3.83	1.00	0.90	0.30	0.23	1.40	
NIST-F1	54844	54859	5.3	0.3	0.3	0.6	0.6	0.9	
NIST-F1	54864	54879	5.8	0.3	0.3	0.5	0.6	0.9	
NIST-F1	54904	54919	4.1	0.3	0.3	0.6	0.6	1.0	
NIST-F1	54924	54939	6.8	0.3	0.3	0.4	0.6	0.9	
NIST-F1	54969	54979	6.9	0.3	0.3	0.3	0.9	1.0	
NIST-F1	54994	55009	5.93	0.30	0.32	0.53	0.61	0.92	
NIST-F1	55114	55134	6.35	0.43	0.31	0.24	0.47	0.75	
NIST-F1	55134	55149	4.20	0.33	0.37	0.28	0.61	0.83	
NIST-F1	55184	55194	4.06	0.56	0.31	0.32	1.05	1.27	
NMIJ-F1	54839	54859	2.6	0.8	3.9	0.3	0.7	4.0	
NMIJ-F1	54859	54889	5.8	0.7	3.9	0.3	0.5	4.0	
NMIJ-F1	54889	54919	6.2	0.7	3.9	0.3	0.5	4.0	
NMIJ-F1	54919	54949	6.2	0.7	3.9	0.3	0.5	4.0	
NMIJ-F1	54949	54979	5.6	0.7	3.9	0.3	0.5	4.0	
NMIJ-F1	54979	55009	5.69	0.70	3.90	0.30	0.46	4.00	
NMIJ-F1	55009	55039	4.82	0.70	3.90	0.30	0.46	4.00	
PTB-CS1	54829	54859	-9.8	5.0	8.0	0.0	0.2	9.4	(1)
PTB-CS1	54859	54889	-13.0	5.0	8.0	0.0	0.2	9.4	
PTB-CS1	54889	54919	-9.6	5.0	8.0	0.0	0.1	9.4	
PTB-CS1	54919	54949	6.9	5.0	8.0	0.0	0.1	9.4	
PTB-CS1	54949	54979	-6.4	5.0	8.0	0.0	0.1	9.4	
PTB-CS1	54979	55009	2.61	5.00	8.00	0.00	0.13	9.43	
PTB-CS1	55009	55039	-8.31	5.00	8.00	0.00	0.13	9.43	
PTB-CS1	55039	55074	-0.80	5.00	8.00	0.00	0.11	9.43	
PTB-CS1	55074	55104	-5.26	5.00	8.00	0.00	0.13	9.43	
PTB-CS1	55104	55134	-11.55	5.00	8.00	0.00	0.13	9.43	
PTB-CS1	55134	55164	-12.08	5.00	8.00	0.00	0.13	9.43	
PTB-CS1	55164	55194	2.04	5.00	8.00	0.00	0.13	9.43	
PTB-CS2	54829	54859	3.7	3.0	12.0	0.0	0.2	12.4	(1)
PTB-CS2	54859	54889	3.6	3.0	12.0	0.0	0.2	12.4	
PTB-CS2	54889	54919	1.4	3.0	12.0	0.0	0.1	12.4	
PTB-CS2	54919	54949	5.5	3.0	12.0	0.0	0.1	12.4	
PTB-CS2	54949	54979	4.3	3.0	12.0	0.0	0.1	12.4	
PTB-CS2	54979	55009	9.28	3.00	12.00	0.00	0.13	12.37	
PTB-CS2	55009	55039	3.03	3.00	12.00	0.00	0.13	12.37	
PTB-CS2	55039	55074	7.17	3.00	12.00	0.00	0.11	12.37	
PTB-CS2	55074	55104	-1.40	3.00	12.00	0.00	0.13	12.37	
PTB-CS2	55104	55134	0.80	3.00	12.00	0.00	0.13	12.37	
PTB-CS2	55134	55164	5.74	3.00	12.00	0.00	0.13	12.37	
PTB-CS2	55164	55194	1.58	3.00	12.00	0.00	0.13	12.37	
PTB-CSF1	54839	54859	5.9	0.1	0.9	0.0	0.3	1.0	
PTB-CSF1	55019	55049	7.64	0.13	0.76	0.01	0.13	0.78	

Table 6. (Cont.)

Standard	Period of estimation		$d/10^{-15}$	$u_A/10^{-15}$	$u_B/10^{-15}$	$u_{\text{link/lab}}/10^{-15}$	$u_{\text{link/TAI}}/10^{-15}$	$u/10^{-15}$	Notes
PTB-CSF2	54819	54829	4.21	0.70	1.06	0.11	0.53	1.38	
PTB-CSF2	54839	54854	4.29	0.70	0.67	0.06	0.37	1.04	
PTB-CSF2	54884	54899	5.35	0.70	0.82	0.30	0.24	1.15	
PTB-CSF2	55029	55049	5.83	0.70	1.07	0.03	0.19	1.29	
PTB-CSF2	55064	55084	6.03	0.70	1.02	0.03	0.19	1.25	
PTB-CSF2	55099	55109	4.24	0.70	1.35	0.03	0.35	1.56	
SYRTE-F01	54829	54839	5.7	0.3	0.4	0.1	0.9	1.0	
SYRTE-F01	54849	54859	5.2	0.2	0.4	0.1	0.9	1.0	
SYRTE-F01	54859	54884	5.1	0.3	0.4	0.1	0.4	0.6	
SYRTE-F01	54894	54919	4.3	0.3	0.4	0.1	0.4	0.6	
SYRTE-F01	54919	54949	4.7	0.3	0.4	0.1	0.3	0.6	
SYRTE-F01	54949	54979	4.6	0.3	0.5	0.1	0.3	0.6	
SYRTE-F01	55054	55074	4.20	0.20	0.43	0.11	0.66	0.82	
SYRTE-F01	55074	55104	4.88	0.20	0.47	0.11	0.40	0.66	
SYRTE-F01	55104	55134	2.63	0.20	0.42	0.11	0.33	0.58	
SYRTE-F01	55134	55164	2.79	0.20	0.41	0.11	0.33	0.57	
SYRTE-F02	54829	54859	6.5	0.3	0.5	0.1	0.3	0.7	
SYRTE-F02	54934	54949	5.1	0.5	0.5	0.1	0.6	0.9	
SYRTE-F02	54969	54979	4.4	0.6	0.5	0.1	0.9	1.2	
SYRTE-F02	55044	55074	5.31	0.50	0.45	0.11	0.46	0.82	
SYRTE-F02	55074	55104	6.05	0.20	0.45	0.11	0.40	0.64	
SYRTE-F02	55104	55134	3.90	0.20	0.45	0.10	0.33	0.60	
SYRTE-F02	55139	55164	3.97	0.20	0.48	0.11	0.38	0.66	
SYRTE-F02	55164	55194	4.54	0.40	0.39	0.11	0.36	0.67	
SYRTE-F0M	54919	54944	6.1	0.2	0.7	2.0	0.4	2.2	(2)
SYRTE-F0M	55044	55074	5.70	0.20	0.71	2.00	0.46	2.18	(2)
SYRTE-F0M	55074	55104	5.07	0.20	0.71	0.35	0.40	0.91	(2)
SYRTE-F0M	55104	55119	2.38	0.20	0.71	0.56	0.61	1.11	(2)
SYRTE-JP0	54839	54859	13.6	0.9	6.3	0.3	0.5	6.4	
SYRTE-JP0	54869	54889	8.1	0.8	6.3	0.3	0.5	6.4	
SYRTE-JP0	54889	54919	5.7	0.8	6.3	0.3	0.3	6.4	
SYRTE-JP0	54919	54949	4.3	0.7	6.3	0.3	0.3	6.4	
SYRTE-JP0	54949	54979	5.2	0.6	6.3	0.3	0.3	6.3	
SYRTE-JP0	54979	55009	5.63	0.65	6.30	0.30	0.33	6.35	
SYRTE-JP0	55009	55039	3.90	0.61	6.30	0.30	0.33	6.34	
SYRTE-JP0	55039	55044	4.59	1.39	6.30	0.30	1.99	6.76	
SYRTE-JP0	55069	55074	1.78	1.44	6.30	0.30	2.29	6.86	
SYRTE-JP0	55074	55104	3.54	0.56	6.30	0.30	0.40	6.34	
SYRTE-JP0	55104	55134	1.31	0.54	6.30	0.30	0.33	6.34	
SYRTE-JP0	55134	55164	1.57	0.53	6.30	0.30	0.33	6.34	
SYRTE-JP0	55164	55194	2.49	0.53	6.30	0.30	0.36	6.34	

**Notes:**

(1) Continuously operating as a clock participating to TAI.

(2) Operated in Toulouse (France)

**References:**

- [1] Levi F. et al., *Metrologia* **43**(6), 545, 2006.
- [2] Lee S.H. et al. , *Metrologia* **46**(3), 227-236, 2009.
- [3] Kumagai M., et al., *Metrologia* **45**(2), 139-148, 2008.
- [4] Heavner T.P. et al., *Metrologia* **42**, 411, 2005. Parker T.E. et al., *Metrologia* **42**, 423, 2005.
- [5] Kurosu T. et al., *IEEE Trans. IM* **53**, 466, 2004.
- [6] Bauch A. et al., *Metrologia* **35**, 829, 1998; Bauch A., *Metrologia* **42**, S43, 2005.
- [7] Bauch A. et al., *IEEE Trans. IM* **36**, 613, 1987; Bauch A., *Metrologia* **42**, S43, 2005.
- [8] Weyers S. et al., *Metrologia* **38**(4), 343, 2001; Weyers S. et al., *Proceedings of the 6th Symposium on Frequency Standards and Metrology, University of St Andrews, World Scientific Pub.*, 64-71, 2001.
- [9] Gerginov V. Et al., *Metrologia* **47**(1), 65-79, 2010.
- [10] Vian C. et al., *IEEE Trans. IM* **54**, 833, 2005 ; Vian C. et al., *Proc 19<sup>th</sup> EFTF*, **53**, 2005.
- [11] Marion H. et al. *Phys. Rev. Lett.* **90**, 150801, 2003.
- [12] Makdissi A. and de Clercq E., *Metrologia* **38**(5), 409, 2001.