

Table 6. Measurements of the duration of the TAI scale interval

(File available at [ftp://62.161.69.5/pub/tai/scale/utai11.ar](http://62.161.69.5/pub/tai/scale/utai11.ar))

TAI is a realization of coordinate time TT. The following tables give the fractional deviation d of the scale interval of TAI from that of TT (in practice the SI second on the geoid), i.e. the fractional frequency deviation of TAI with the opposite sign: $d = -y_{\text{TAI}}$.

In this table, d is obtained on the given periods of estimation by comparison of the TAI frequency with that of the individual primary frequency standards (PFS) IT-CSF1, NICT-CSF1, NIST-F1, NMIJ-F1, NPL-CSF2, PTB-CS1, PTB-CS2, PTB-CSF1, PTB-CSF2, SYRTE-FO1, SYRTE-FO2, SYRTE-FOM and SYRTE-JPO for the year 2011. Previous calibrations are available in the successive annual reports of the BIPM Time Section volumes 1 to 18 and in the BIPM annual report on time activities volumes 1 to 5.

Each comparison is provided with the following information:

u_A is the uncertainty originating in the instability of the PFS,

u_B is the combined uncertainty from systematic effects,

$u_{\text{link/lab}}$ is the uncertainty in the link between the PFS and the clock participating to TAI, including the uncertainty due to dead-time,

$u_{\text{link/TAI}}$ is the uncertainty in the link to TAI, computed using the standard uncertainty of [UTC-UTC(k)],

u is the quadratic sum of all four uncertainty values.

In this table, a frequency over a time interval is defined as the ratio of the end-point phase difference to the duration of the interval.

The typical characteristics of the calibrations of the TAI frequency provided by the different primary standards over 2011 are indicated below. Reports of individual PFS evaluations may be found at [ftp://62.161.69.5/pub/tai/data/PFS_reports](http://62.161.69.5/pub/tai/data/PFS_reports). Subdirectory named 'data'. Ref(u_B) is a reference giving information on the stated value of u_B , $u_B(\text{Ref})$ is the u_B value stated in this reference. Note that the current u_B values are generally not the same as the peer reviewed values given in Ref(u_B).

Primary Standard	Type /selection	Type B std. uncertainty	$u_B(\text{Ref})/10^{-15}$	Ref(u_B)	Comparison with	Number/typical duration of comp.
IT-CSF1	Fountain	0.7×10^{-15}	0.5	[1]	H maser	1 / 25 d
NICT-CSF1	Fountain	$(1.0 \text{ to } 1.2) \times 10^{-15}$	1.9	[2]	UTC(NICT)	2 / 10 d to 20 d
NIST-F1	Fountain	0.31×10^{-15}	0.35	[3]	H maser	5 / 15 d to 30 d
NMIJ-F1	Fountain	3.9×10^{-15}	3.9	[4]	H maser	2 / 30 d
NPL-CSF2	Fountain	$(0.40 \text{ or } 0.41) \times 10^{-15}$	0.41	[5]	H maser	5 / 15 d to 25 d
		0.23×10^{-15}	0.23	[6]	H maser	2 / 15 d to 25 d
PTB-CS1	Beam /Mag.	8×10^{-15}	8.	[7]	TAI	12 / 30 d to 35 d
PTB-CS2	Beam /Mag.	12×10^{-15}	12.	[8]	TAI	7 / 30 d to 35 d
PTB-CSF1	Fountain	$(0.74 \text{ to } 0.79) \times 10^{-15}$	1.4	[9]	H maser	10 / 15 d to 25 d
PTB-CSF2	Fountain	$(0.36 \text{ to } 0.54) \times 10^{-15}$	0.8	[10]	H maser	5 / 15 d to 25 d
		0.56×10^{-15}	0.41	[11]		1 / 20 d
SYRTE-FO1	Fountain	$(0.42 \text{ to } 0.49) \times 10^{-15}$	0.72	[12]	H maser	6 / 10 d to 25 d
SYRTE-FO2	Fountain	$(0.26 \text{ to } 0.39) \times 10^{-15}$	0.65	[12]	H maser	12 / 15 d to 35 d
SYRTE-FOM	Fountain	$(0.82 \text{ to } 0.92) \times 10^{-15}$	0.80	[13]	H maser	6 / 20 d to 30 d

More detailed information on the characteristics and operation of individual PFS may be found in the annexes supplied by the individual laboratories.

Table 6. (Cont.)

Standard	Period of estimation		$d/10^{-15}$	$u_A/10^{-15}$	$u_B/10^{-15}$	$u_{\text{link/lab}}/10^{-15}$	$u_{\text{link/TAI}}/10^{-15}$	$u/10^{-15}$	Notes
IT-CsF1	55564	55589	9.91	0.30	0.70	0.30	0.54	0.98	
NICT-CsF1	55554	55574	3.79	1.00	1.00	0.30	0.28	1.47	
NICT-CsF1	55864	55874	4.15	1.00	1.20	0.30	0.53	1.68	
NIST-F1	55574	55589	5.20	0.46	0.31	0.25	0.37	0.71	
NIST-F1	55634	55649	5.71	0.50	0.31	0.18	0.37	0.72	
NIST-F1	55744	55759	7.25	0.60	0.31	0.29	0.37	0.82	
NIST-F1	55794	55814	6.89	0.44	0.31	0.19	0.28	0.64	
NIST-F1	55894	55924	2.38	0.31	0.31	0.20	0.20	0.52	
NMIJ-F1	55559	55589	6.34	0.70	3.90	0.30	0.20	3.98	
NMIJ-F1	55589	55619	5.92	0.70	3.90	0.30	0.20	3.98	
NPL-CsF2	55554	55569	8.10	0.29	0.40	0.10	0.97	1.10	
NPL-CsF2	55579	55599	5.13	0.23	0.40	0.01	0.75	0.88	
NPL-CsF2	55669	55684	4.89	0.32	0.41	0.07	0.97	1.11	
NPL-CsF2	55709	55734	3.78	0.21	0.41	0.03	0.46	0.65	
NPL-CsF2	55744	55759	4.52	0.27	0.40	0.03	0.49	0.69	
NPL-CsF2	55789	55804	4.47	0.28	0.23	0.01	0.49	0.61	
NPL-CsF2	55864	55889	3.88	0.22	0.23	0.02	0.23	0.39	
PTB-CS1	55559	55589	-2.21	6.00	8.00	0.00	0.13	10.00	(1)
PTB-CS1	55589	55619	4.11	6.00	8.00	0.00	0.13	10.00	
PTB-CS1	55619	55649	-6.78	6.00	8.00	0.00	0.13	10.00	
PTB-CS1	55649	55679	-11.60	6.00	8.00	0.00	0.13	10.00	
PTB-CS1	55679	55709	-3.20	6.00	8.00	0.00	0.13	10.00	
PTB-CS1	55709	55739	-9.15	6.00	8.00	0.00	0.13	10.00	
PTB-CS1	55739	55769	-8.61	6.00	8.00	0.00	0.13	10.00	
PTB-CS1	55769	55804	-2.04	6.00	8.00	0.00	0.11	10.00	
PTB-CS1	55804	55834	-5.33	6.00	8.00	0.00	0.13	10.00	
PTB-CS1	55834	55864	-8.65	6.00	8.00	0.00	0.13	10.00	
PTB-CS1	55864	55894	-6.68	6.00	8.00	0.00	0.13	10.00	
PTB-CS1	55894	55924	-12.00	6.00	8.00	0.00	0.13	10.00	
PTB-CS2	55709	55739	1.31	3.00	12.00	0.00	0.13	12.37	(1)
PTB-CS2	55739	55769	9.06	3.00	12.00	0.00	0.13	12.37	
PTB-CS2	55769	55804	1.47	3.00	12.00	0.00	0.11	12.37	
PTB-CS2	55804	55834	1.31	3.00	12.00	0.00	0.13	12.37	
PTB-CS2	55834	55864	-0.97	3.00	12.00	0.00	0.13	12.37	
PTB-CS2	55864	55894	5.36	3.00	12.00	0.00	0.13	12.37	
PTB-CS2	55894	55924	5.90	3.00	12.00	0.00	0.13	12.37	
PTB-CSF1	55539	55564	7.03	0.23	0.76	0.02	0.15	0.81	
PTB-CSF1	55569	55589	6.48	0.32	0.76	0.03	0.19	0.85	
PTB-CSF1	55634	55654	5.68	0.11	0.76	0.02	0.19	0.79	
PTB-CSF1	55659	55674	7.04	0.14	0.76	0.02	0.24	0.81	
PTB-CSF1	55694	55709	6.92	0.24	0.76	0.02	0.24	0.83	
PTB-CSF1	55714	55729	6.64	0.17	0.79	0.02	0.24	0.84	
PTB-CSF1	55774	55794	5.79	0.12	0.77	0.01	0.19	0.80	
PTB-CSF1	55804	55824	4.67	0.12	0.74	0.01	0.19	0.77	
PTB-CSF1	55854	55869	4.75	0.15	0.74	0.01	0.24	0.79	
PTB-CSF1	55874	55894	4.49	0.15	0.74	0.01	0.19	0.78	

Table 6. (Cont.)

Standard	Period of estimation		$d/10^{-15}$	$u_A/10^{-15}$	$u_B/10^{-15}$	$u_{\text{link/lab}}/10^{-15}$	$u_{\text{link/TAI}}/10^{-15}$	$u/10^{-15}$	Notes
PTB-CSF2	55629	55644	6.81	0.21	0.54	0.02	0.24	0.63	
PTB-CSF2	55659	55679	7.31	0.16	0.46	0.02	0.19	0.52	
PTB-CSF2	55694	55709	6.83	0.20	0.42	0.03	0.24	0.53	
PTB-CSF2	55774	55799	5.69	0.20	0.43	0.04	0.15	0.50	
PTB-CSF2	55824	55839	4.33	0.28	0.36	0.04	0.24	0.52	
PTB-CSF2	55859	55879	4.24	0.17	0.56	0.06	0.19	0.62	
SYRTE-F01	55639	55649	4.17	0.30	0.45	0.10	0.53	0.76	
SYRTE-F01	55664	55679	5.17	0.20	0.49	0.14	0.37	0.66	
SYRTE-F01	55684	55709	5.05	0.40	0.42	0.14	0.23	0.64	
SYRTE-F01	55709	55729	6.36	0.30	0.43	0.13	0.28	0.61	
SYRTE-F01	55869	55889	2.96	0.25	0.43	0.11	0.28	0.58	
SYRTE-F01	55914	55924	1.98	0.20	0.46	0.18	0.53	0.75	
SYRTE-F02	55619	55639	6.10	0.20	0.39	0.11	0.51	0.68	
SYRTE-F02	55594	55619	6.66	0.25	0.39	0.11	0.54	0.72	
SYRTE-F02	55654	55679	5.83	0.20	0.39	0.11	0.23	0.51	
SYRTE-F02	55684	55709	5.68	0.30	0.26	0.13	0.23	0.48	
SYRTE-F02	55709	55734	7.23	0.30	0.26	0.12	0.23	0.47	
SYRTE-F02	55739	55769	6.25	0.40	0.26	0.11	0.20	0.53	
SYRTE-F02	55769	55804	5.08	0.30	0.28	0.13	0.17	0.46	
SYRTE-F02	55804	55819	6.18	0.20	0.27	0.17	0.37	0.52	
SYRTE-F02	55819	55834	3.88	0.30	0.27	0.10	0.37	0.55	
SYRTE-F02	55834	55864	4.79	0.30	0.28	0.14	0.20	0.48	
SYRTE-F02	55864	55894	4.34	0.40	0.28	0.11	0.20	0.54	
SYRTE-F02	55894	55924	3.22	0.20	0.26	0.11	0.20	0.40	
SYRTE-F0M	55594	55614	6.52	0.30	0.82	0.22	0.66	1.12	
SYRTE-F0M	55629	55649	4.66	0.20	0.92	0.22	0.28	1.01	
SYRTE-F0M	55649	55679	5.77	0.30	0.82	0.37	0.20	0.97	
SYRTE-F0M	55684	55709	5.79	0.20	0.82	0.12	0.23	0.88	
SYRTE-F0M	55869	55894	3.86	0.25	0.82	0.11	0.23	0.89	
SYRTE-F0M	55894	55924	2.94	0.20	0.82	0.12	0.20	0.87	

Notes:

(1) Continuously operating as a clock participating to TAI.

References:

- [1] Levi F. et al., [Metrologia 43\(6\), 545, 2006.](#)
- [2] Kumagai M., et al., [Metrologia 45\(2\), 139-148, 2008.](#)
- [3] Heavner T.P. et al., [Metrologia 42, 411, 2005.](#) Parker T.E. et al., [Metrologia 42, 423, 2005.](#)
- [4] Kurosu T. et al., *IEEE Trans. IM* **53**, 466, 2004.
- [5] Szymaniec K., Park S.E., Marra G. and Chalupczak W., *Metrologia* **47(4)**, 363-376, 2010.
- [6] R. Li, K. Gibble and K. Szymaniec, *Metrologia*, 48, pp. 283-289, 2011.
- [7] Bauch A. et al., [Metrologia 35, 829, 1998;](#) Bauch A., [Metrologia 42, S43, 2005.](#)
- [8] Bauch A. et al., *IEEE Trans. IM* **36**, 613, 1987; Bauch A., [Metrologia 42, S43, 2005.](#)
- [9] Weyers S. et al., [Metrologia 38\(4\), 343, 2001;](#) Weyers S. et al., *Proceedings of the 6th Symposium on Frequency Standards and Metrology, University of St Andrews, World Scientific Pub.*, 64-71, 2001.
- [10] Gerginov V. Et al., [Metrologia 47\(1\), 65-79, 2010.](#)
- [11] S. Weyers, V. Gerginov, N. Nemitz, R. Li and K. Gibble, *Metrologia* 49, 82-87 (2012).
- [12] Vian C. et al., *IEEE Trans. IM* **54**, 833, 2005 ; Vian C. et al., *Proc 19th EFTF*, **53**, 2005.
- [13] Marion H. et al. *Phys. Rev. Lett.* **90**, 150801, 2003.