

Annex C: Uncertainty budget terms

1 Statistical uncertainty

The statistical uncertainty $u_a(A-B)$ for the comparison between two receivers A and B is evaluated by taking the upper limit of the error bar of the TDEV at 61 440s (close to 1 d) when available, or otherwise the upper limit of the last error bar available. When two traveling receivers are used for the campaign the retained value is the average of the two statistical uncertainty evaluated as above. The statistical uncertainty of the calibration is obtained by the quadratic sum of the uncertainty at reference site and the one at visited site.

$$u_a^2 = u_a^2(T-R) - u_a^2(T-V)$$

2 Type B uncertainties

- $u_{b,1}$ observed maximum misclosure. This uncertainty component is an estimation of the stability of equipment during the campaign. When a single receiver is traveling the only possible estimation is the classical misclosure. With two traveling receivers it is also possible to estimate the stability of the ensemble during the trip by computing the offset between these receivers when implemented in all sites. The misclosure $u_{b,1}$ we use is the maximum value between the actual misclosure between the start and the end of the campaign and the offset between both traveling equipment.
- $u_{b,11}$ position error at reference site. The conventional value agreed at CCTF WG on GNSS is 200 ps.
- $u_{b,12}$ position error at visited site. The conventional value agreed at CCTF WG on GNSS is 200 ps.
- $u_{b,13}$ multipath at reference site. We propose a value of 50 ps. It is smaller than the conventional value proposed by the CCTF WG on GNSS, but it is in line with some preliminary experiment at OP.
- $u_{b,14}$ multipath at visited site. Same as above.
- $u_{b,21}$ REFDFLY (traveling receiver at reference lab). Uncertainty of the measure of the time difference between the reference point of the traveling receiver and the local UTC(k). The used value is the quadratic sum of a conventional value (200 ps) with the standard deviation of the actual measurement. When the REFDFLY is obtained by summing several individual measurement the uncertainty is increased by quadratic sum as required.
- $u_{b,22}$ REFDFLY (traveling receiver at visited lab). Same as above.
- $u_{b,TOT}$: Quadratic sum of all previous u_b .
- $u_{b,31}$ REFDFLY uncertainty of the reference system to its local UTC(k). Computed as $u_{b,21}$. This term can be set to 0 when the reference has been recently calibrated, the uncertainty of REFDFLY being already included in the calibration of the reference receiver.
- $u_{b,32}$ REFDFLY uncertainty (at visited lab) of the link of the visited system to its local UTC(k). Computed as $u_{b,21}$. When this delay is measured and the $u_{b,32}$ is taken into

account, the local distribution system can be modified afterwards without losing the calibration.

- $u_{b,41}$ uncertainty of Antenna Cable delay at reference station. Same consideration as for REFDFLY above.
- $u_{b,42}$ uncertainty of Antenna Cable delay at visited station. Same consideration as for REFDFLY above. When for some reason the Antenna Cable of the traveling system is changed during the campaign $u_{b,42}$ is obtained by the quadratic sum of the uncertainty of the Antenna Cable of the visited station and of the traveling equipment.
- $u_{b,SYS}$: Quadratic sum of all type B uncertainty.
- u_{CAL0} : Quadratic sum of u_a and $u_{b,SYS}$. This uncertainty is for the link between the calibrated receiver and the reference receiver, without taking into account the uncertainty of this reference receiver.