

Calibrations between PTB and OP

d are differential time corrections to be added to [$UTC(PTB) - UTC(OP)$],
and $u(d)$ are estimated uncertainties for the periods of comparisons.

Date	<i>d</i> /ns	<i>u</i> (<i>d</i>)/ns	Reference	Note
November 2009	311.4	3.0	BIPM Report-2010/04	(1)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the ‘round-trip’ reproducibility at the OP.

Notes

- (1) TTS-3 receiver (s/n: 0014) at PTB