

Calibrations between KRIS and OP

d are differential time corrections to be added to $[UTC(KRIS) - UTC(OP)]$,
and $u(d)$ are estimated uncertainties for the periods of comparisons.

Date	d/ns	$u(d)/ns$	Reference	Note
October 2003	8.8	3.0	BIPM Report-2004/06	(1)
December 2004	-2.6	3.0	BIPM Report-2008/02	(2)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the 'round-trip' reproducibility at the OP.

Notes

- (1) Topcon Euro-80 receiver (s/n: 023C10474) at KRIS
- (2) Euro-80 receiver (s/n: IRT0112301) at KRIS