

Calibrations between MKEH (*) and OP

d are differential time corrections to be added to [$UTC(MKEH) - UTC(OP)$],
and $u(d)$ are estimated uncertainties for the periods of comparisons.

Date	<i>d</i> /ns	$u(d)$ /ns	Reference	Note
November 2006	54.5	3.0	BIPM Report-2010/02	(1)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the ‘round-trip’ reproducibility at the OP.

Notes

(1) TTR-6 receiver (s/n: 028) at MKEH

(*) MKEH: formerly OMH.