

Calibrations between NMIJ and OP

d are differential time corrections to be added to $[UTC(NMIJ) - UTC(OP)]$,
and $u(d)$ are estimated uncertainties for the periods of comparisons.

Date	d/ns	$u(d)/ns$	Reference	Note
September 2002	-20.0	4.0	BIPM Report-2003/05	(1)
January 2005	-16.8	4.0	BIPM Report-2008/02	(1)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the 'round-trip' reproducibility at the OP.

Notes

- (1) TTR-6 receiver (s/n: 484) at NMIJ