

## Calibrations between PL (\*) and OP

$d$  are differential time corrections to be added to  $[UTC(PL) - UTC(OP)]$ ,  
and  $u(d)$  are estimated uncertainties for the periods of comparisons.

Date	$d/ns$	$u(d)/ns$	Reference	Note
April 1995	-5.0	2.0	Rapport BIPM-1995/11	(1)
October 2006	-3.4	3.0	<a href="#">BIPM Report-2010/02</a>	(2)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the 'round-trip' reproducibility at the OP.

### Notes

- (1) TTR-6 receiver (s/n: 282) at GUM
- (2) TTS-2 receiver (s/n: 014) at GUM
  
- (\*) UTC(PL) is realized at GUM.