## Calibrations between ROA and OP

d are differential time corrections to be added to [UTC(ROA) - UTC(OP)], and u(d) are estimated uncertainties for the periods of comparisons.

Date	d/ns	u(d)/ns	Reference	Note
April 1998	-8.0	2.0	Rapport BIPM-1998/07	(1)
June 1998	-6.0	3.0	[1]	
March 2002	-2.0	3.0	BIPM Report-2002/02	(1)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the 'round-trip' reproducibility at the OP.

## References

[1] Lewandowski, Konate, 4th trip

## **Notes**

(1) TTR-6 receiver (s/n: 253) at ROA