

## Calibrations between SCL and OP

*d* are differential time corrections to be added to [ $UTC(SCL) - UTC(OP)$ ],  
and  $u(d)$  are estimated uncertainties for the periods of comparisons.

Date	<i>d</i> /ns	<i>u(d)</i> /ns	Reference	Note
May 1993	1.8	1.5	Rapport BIPM-1993/05	(1)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the ‘round-trip’ reproducibility at the OP.

### Notes

- (1) TTR-5A receiver (s/n: 0192) at SCL