

Calibrations between SP and OP

d are differential time corrections to be added to $[UTC(SP) - UTC(OP)]$,
and $u(d)$ are estimated uncertainties for the periods of comparisons.

Date	d/ns	$u(d)/ns$	Reference	Note
October 1997	-43.0	3.0	Rapport BIPM-1998/01	(1)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the 'round-trip' reproducibility at the OP.

Notes

- (1) TTR-6 receiver (s/n: 425) at SP