

Calibrations between TL and OP

d are differential time corrections to be added to $[UTC(TL) - UTC(OP)]$,
and $u(d)$ are estimated uncertainties for the periods of comparisons.

Date	d/ns	$u(d)/ns$	Reference	Note
September 2002	-54.0	4.0	BIPM Report-2003/05	(1)
October 2004	-6.8	3.0	BIPM Report-2008/02	(2)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the 'round-trip' reproducibility at the OP.

Notes

- (1) TTR-6 receiver (s/n: 479) at TL
- (2) Topcon Euro-80 receiver at TL