

# GNSS calibration of INRIM and RISE receivers with respect to OP G1 (1018-2022)

## Summary

In August 2022, the Laboratoire national de métrologie et d'essais, Systèmes de références Temps-Espace, Observatoire de Paris, Observatoire de Paris (OP) conducted a trip to calibrate GNSS equipment owned by the Istituto Nazionale di Ricerca Metrologica (INRIM) and the Research Institutes of Sweden AB (RISE). The trip started and finished at the OP, providing closure with respect to the OP Group1 reference receiver OP73.

The operations and report of measurements are described in the [report by OP](#).

- **Final results for the calibrated systems**

The INTDLY or SYSDLY values of the receivers given in Table 1 and Table 2 have been computed by OP based on the results of the Group 1 trip [1001-2020](#) for OP73 (GPS and Galileo) and should not be updated to reflect later changes in the conventional INTDLY values of the reference receiver.

For a P3/E3/PPP UTC link A-B involving any Group 1 and any receiver in this trip, the uncertainty resulting from the calibration,  $U_B(A-B)$ , is computed as

$$U_B(A-B) = (U_{CAL0}^2 + \Delta U_{CAL}(A)^2 + \Delta U_{CAL}(B)^2)^{1/2} \quad (1)$$

where  $U_{CAL0} = 2.5$  ns is the conventional Group 2 value, and where  $\Delta U_{CAL}$  (generally zero) is specified for each system.

Changes in the set-up of the receivers after the calibration must be accounted for as described in section A.3.6 of the most recent Calibration guidelines in <https://webtai.bipm.org/ftp/pub/tai/publication/gnss-calibration/guidelines/>.

Table 1. Final P1/P2 INTDLY values from the 1018-2022 exercise. Values of REFDLY and CABDLY used to compute calibration results are also indicated for reference. All values are in ns. “Meas. Date” refers to the first day of the differential calibration, to which the calibration results can be applied. “Impl. Date” is the MJD when the results should be implemented in the receiver.

System	BIPM	Meas. date	INTDLY P1	INTDLY P2	REF DLY	CAB DLY	Note	$\Delta U_{CAL}$	Impl. date
GR01	IT11	2022/08/26	56.0	54.4	483.6	298.6		0.0	60095
GR02	IT12	2022/08/26	352.8	352.7	483.5	0.0	(1)	0.0	60095
GR04	IT14	2022/08/26	279.8	277.5	376.2	0.0	(1)	0.0	60095
GR05	IT15	2022/08/26	27.5	24.4	432.9	306.4		0.0	60095
GR06	IT16	2022/08/26	27.6	25.5	161.1	355.5		0.0	60095
INR5	IT08	2022/08/26	309.0	308.5	483.9	0.0	(1)	0.0	60095
INR6	IT09	2022/08/26	55.0	55.5	484.0	130.5		0.0	60095
GTRB	IT2_	2022/08/26	-37.1	-23.0	107.3	131.0		0.0	60095
IENG	IT10	2022/08/26	54.3	55.0	483.6	130.5		0.0	60095
SP03	SP03	2022/10/28	268.1	265.0	0.0	0.0	(2)	0.0	60095
SP05	SP05	2022/10/28	200.7	195.7	0.0	0.0	(2)	0.0	60095
RIT1	SP06	2022/10/28	273.7	269.7	0.0	0.0	(2)	0.0	60095
RIT2	SP07	2022/10/28	231.7	228.2	0.0	0.0	(2)	0.0	60095

## Notes:

- (1) Results expressed in terms of SYSDLY
- (2) Results expressed in terms of TOTDLY

Table 2. Final E1/E5a INTDLY values from the 1018-2022 exercise. Values of REFDLY and CABDLY used to compute calibration results are also indicated for reference. All values are in ns. "Meas. Date" refers to the first day of the differential calibration, to which the calibration results can be applied. "Impl. Date" is the MJD when the results should be implemented in the receiver.

System	BIPM	Meas. date	INTDLY E1	INTDLY E5a	REF DLY	CAB DLY	Note	$\Delta U_{CAL}$	Impl. date
GR01	IT11	2022/08/26	56.9	63.9	483.6	298.6	(3)	1.0	60095
GR02	IT12	2022/08/26	353.7	362.4	483.5	0.0	(1) (3)	1.0	60095
GR04	IT14	2022/08/26	282.1	281.9	376.2	0.0	(1) (3)	1.0	60095
GR05	IT15	2022/08/26	29.7	28.4	432.9	306.4	(3)	1.0	60095
GR06	IT16	2022/08/26	29.9	28.3	161.1	355.5	(3)	1.0	60095
INR5	IT08	2022/08/26	309.9	320.3	483.9	0.0	(1) (3)	1.0	60095
INR6	IT09	2022/08/26	55.8	65.1	484.0	130.5	(3)	1.0	60095
IENG	IT10	2022/08/26	55.1	64.5	483.6	130.5	(3)	1.0	60095
SP03	SP03	2022/10/28	270.5	282.9	0.0	0.0	(2) (3)	1.0	60095
SP05	SP05	2022/10/28	203.3	202.1	0.0	0.0	(2) (3)	1.0	60095
RIT1	SP06	2022/10/28	276.1	273.5	0.0	0.0	(2) (3)	1.0	60095
RIT2	SP07	2022/10/28	234.6	238.1	0.0	0.0	(2) (3)	1.0	60095

## Notes:

- (1) Results expressed in terms of SYSDLY
- (2) Results expressed in terms of TOTDLY
- (3) Extra uncertainty due to misclosure values larger than usual

## Version history

V1.0 2023/03/22: Publication of results from V1.1 of the OP report.

V1.1 2023/07/12: Date typos in the text corrected